



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product **28055 - Sabouraud Dextrose Agar**
Batch No **1404888**
Analysis Date 07-Oct-2024
Date of Manufacture October 2024
Expiry Date/Re-Test Date October 2027

Test Parameters	Standards	Actual Results
Appearance (Colour)	Light biege	Complies
Appearance (Form)	Free flowing, homogeneous powder	Complies
Solubility	65.00 gm/liter	Complies
Solubility before autoclaving (Clarity)	Very slightly opalescent	Complies
Gel strength (1.5% gel)	Firm, comparable with 1.5% gel	Complies
pH (25°C)	5.6 ± 0.2	5.70, Complies
Prepared Medium Appearance after autoclaving (Clarity)	Very slightly opalescent	Complies
Prepared Medium Appearance after autoclaving (Colour)	Light amber	Complies
Cultural Response	Inoculate and incubate at 22 - 25°C for 18-48 hours or upto 7 days with following organisms:	
Organism	Candida albicans ATCC 10231	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Aspergillus niger ATCC 16404	
Inoculum (cfu) 10-100	Growth : Good	Complies
Organism	Saccharomyces cerevisiae ATCC 9763	
Inoculum (cfu) 10-100	Growth : Good	Complies

FRM/82/11/3/28055 Rev No. 2 Date 03/07/2021

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.