



**SISCO  
RESEARCH  
LABORATORIES  
PVT. LTD.**

[www.srlchem.com](http://www.srlchem.com)

## Certificate of Analysis

**Product** **19440 - Chromed Chromogenic Coliform Agar w/ SLS (CCA w/ SLS)**  
**Batch No** **2009016**  
**Analysis Date** 11-Nov-2023  
**Date of Manufacture** November 2023  
**Expiry Date/Re-Test Date** November 2025

Test Parameters	Standards	Actual Results
Appearance (Colour)	Cream	Complies
Appearance (Form)	Free flowing, homogeneous powder	Complies
Solubility	31.32 gm/liter	Complies
Solubility before boiling (clarity)	Clear to very slightly opalescent	Complies
Gel strength (1.5% gel)	Firm, comparable with 1.5% agar gel	Complies
pH (25°C)	6.8 ± 0.2	6.73, Complies
Prepared medium appearance after boiling (clarity)	Clear to very slightly opalescent	Complies
Prepared medium appearance after boiling (colour)	Cream to yellow	Complies
Cultural Response	Inoculate and incubate at 35-37°C for 18-24 hours	
Organism	Escherichia coli ATCC 25922	
Inoculum (cfu) 10-100	Growth : Good Recovery rate : >=70% Colony colour : Dark blue to violet	Complies
Organism	Escherichia coli ATCC 8739	
Inoculum (cfu) 10-100	Growth : Good Recovery rate : >=70% Colony colour : Dark blue to violet	Complies
Organism	Enterobacter aerogenes ATCC 13048	
Inoculum (cfu) 10-100	Growth : Good Recovery rate : >=70% Colony colour : Pale pink to pink	Complies
Organism	Pseudomonas aeruginosa ATCC 10145	
Inoculum (cfu) 10-100	Growth : Good Recovery rate : >=70% Colony colour : Colourless	Complies
Organism	Citrobacter freundii ATCC 43864	

FRM/82/11/3/19440 Rev No. 1 Date 24/07/2021

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,  
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/2

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.



**SISCO  
RESEARCH  
LABORATORIES  
PVT. LTD.**

[www.srlchem.com](http://www.srlchem.com)

## Certificate of Analysis

**Product**

**19440 - Chromed Chromogenic Coliform Agar w/ SLS  
(CCA w/ SLS)**

**Batch No**

**2009016**

**Analysis Date**

11-Nov-2023

**Date of Manufacture**

November 2023

**Expiry Date/Re-Test Date**

November 2025

Test Parameters	Standards	Actual Results
Inoculum (cfu) 10-100	Growth : Good Recovery rate : $\geq 70\%$ Colony colour : Pale pink to pink	Complies
Organism	Enterococcus faecalis ATCC 19433	
Inoculum	$\geq 10^3$ , Growth : Inhibited	Complies

FRM/82/11/3/19440 Rev No. 1 Date 24/07/2021

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,  
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

2/2

This COA has been generated electronically and it is valid without signature.

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)