



Certificate of Analysis

Product	10453 - Plate Count Agar
Batch No	7582285
Analysis Date	30-Nov-2023
Date of Manufacture	November 2023
Expiry Date/Re-Test Date	November 2026

Test Parameters	Standards	Actual Results
Appearance (Colour)	Light cream.	Complies
Appearance (Form)	Free flowing, homogeneous powder.	Complies
Solubility	17.50 gm/liter	Complies
Solubility before autoclaving (Clarity)	Clear to very slightly opalescent.	Complies
Gel strength (1.5% gel)	Firm, comparable with 1.5 % gel.	Complies
pH (25°C)	7.0 ± 0.2	7.01, Complies
Prepared Medium Appearance after autoclaving (Clarity)	Clear to very slightly opalescent.	Complies
Prepared Medium Appearance after autoclaving (Colour)	Light amber.	Complies
Cultural Response	Inoculate and incubate at 37°C for 18-24 hours.	
Organism	Staphylococcus aureus ATCC 25923	
Inoculum (cfu) 10-100	Growth : Good,Recovery Rate : >=70%	Complies
Organism	Enterococcus faecalis ATCC 29212	
Inoculum (cfu) 10-100	Growth : Good,Recovery Rate : >=70%	Complies
Organism	Bacillus subtilis ATCC 6633	
Inoculum (cfu) 10-100	Growth : Good,Recovery Rate : >=70%	Complies
Organism	Streptococcus pyogenes ATCC 19615	
Inoculum (cfu) 10-100	Growth : Good,Recovery Rate : >=70%	Complies
Organism	Escherichia coli ATCC 25922	
Inoculum (cfu) 10-100	Growth : Good,Recovery Rate : >=70%	Complies
Organism	Lactobacillus casei ATCC 9595	
Inoculum (cfu) 10-100	Growth : Good,Recovery Rate : >=70%	Complies
Organism	Lactobacillus acidophilus ATCC 11506	
Inoculum (cfu) 10-100	Growth : Good,Recovery Rate : >=70%	Complies

FRM/82/11/3/10453 Rev No. 2 Date 25/06/2021

Prepared by Shikha

In compliance with the standard specifications of SRL

Plant Site 1 : D-88/2, MIDC, Turbhe – 400 705, New Mumbai, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.